

Please find attached modification in Advanced Geotechnical lab invitation.

Following are the modification/addition in the mentioned item number in Annexure 1 of the invitation letter for

Package Code: TEQIP-III/2019/UK/wits/165

Package Name: WIT/CE New/002

Last date of receiving quotations has been extended till 10th February 2020, till 4:30PM

Item No.	Item	Shortcomings
1.	Direct Shear Test App.	Should be supplied with Proving ring & Dial Gauge viz, Proving Ring Capacity 2 kN & Dial Gauge, 0.01 x 25 mm (Set of 2).
2.	Permeability Apparatus	Overhead Tank : should be supplied with appropriate stand and Compaction rammer 2.6 kg.
5.	Laboratory CBR App.	Should be supplied with Proving ring & Dial Gauge viz, Proving Rings for above, 10 kN capacity & 30 kN capacity & Dial Gauge 0.01 x 25mm
7.	Triaxial Shear Test App.	(a) Load Frame - Should be supplied with Proving ring & Dial Gauge viz, Proving Ring Capacity 2 kN & Dial Gauge, 0.01 x 25 mm.
9.	Relative Density App.	Should be supplied with Dial Gauges viz. Dial Gauge 0.01 x 25 mm & 0.01 x 50 mm